



Specification

1.51456.0001 Purospher® STAR RP-18 endcapped (5 µm) Hibar® RT
250-4.6 HPLC column

Spec. Values

Pressure	≤ 130	bar
Theoretical Plates (N/m) (Toluene)	≥ 80000	
Symmetry (Toluene)	0.8 - 1.4	
Capacity factor (Toluene)	1.0 - 3.0	

Peter Schaub

responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature